

कार्यालय महानिदेशक, चिकित्सा एवं स्वास्थ्य सेवायें, उत्तर प्रदेश, लखनऊ।

(सी०एम०एस०डी०, अनुभाग-८)

Phone no: - 0522-2625809

E-mail id: -cmsd.dghealth@gmail.com

संख्या:-८फ/क्यू०सी०-८३०/

लखनऊ: दिनांक: 14 दिसम्बर, 2016

संशोधित निविदा सूचना

महानिदेशालय पत्र संख्या-८फ/क्यू०सी०-८३०/७२६१, दिनांक ०८.११.२०१६ द्वारा ट्रामा सेन्ट्रों, १०० शैय्यायुक्त चिकित्सालयों एवं डॉ० श्यामा प्रसाद मुखर्जी चिकित्सालय, लखनऊ में उपलब्ध ए०एल०एस० एम्बुलेन्स हेतु उपकरणों की आपूर्ति/स्थापना कराये जाने हेतु निविदायें आमंत्रित की गयी थी।

उक्त निविदा के मानकों को निम्नवत् संशोधन किये जाते हैं :-

| Page No. | Existing Specification  | Revised (To be Read as)   |
|----------|---|---|
| 32       | <b>Defibrillator with Monitor</b><br>1.The machine should have facility for ECG monitoring, defibrillation, transcutaneous pacing, defibrillation and synchronized cardio version.<br>5.The machine should have defibrillator facility for neonatal and pediatric patients.<br>6.The machine should have ECG waveform display with provision for synchronization. | <b>Defibrillator with Monitor (NIBP, ECG, SPO2 with AED)</b><br>1.The machine should have facility for ECG monitoring, defibrillation, NIBP, SpO2, AED, transcutaneous pacing, defibrillation and synchronized cardio version.<br>5.The machine should have defibrillator facility for Adult and pediatric patients.<br>6.The machine should have at least 7" or more color screen to display ECG waveform, Spo2 graph, NIBP data with provision for synchronization. |
| 46       | <b>Portable USG Machine</b><br>17. Should be USFDA approved or European CE (Notified Body) approved.  | <b>Portable USG Machine</b><br>17.Should be USFDA approved with 510 K number.   |


निविदा की शेष नियम व शर्तें यथावत् रहेंगी।

पृष्ठांकन संख्या:-८फ/क्यू०सी०-८३०/०२९०-९२

निदेशक (भण्डार)  
तद्दिनांक।

प्रतिलिपि निम्नलिखित को सूचनार्थ एवं आवश्यक कार्यवाही हेतु प्रेषित :-

1. डॉ० अवधेश रावत, एन.आई.सी. उ०प्र० लखनऊ को इस आशय से प्रेषित कि निविदा की सूचना को वेबसाइट :-  
[www.dghealth.up.nic.in](http://www.dghealth.up.nic.in) एवं [www.uphealth.up.nic.in](http://www.uphealth.up.nic.in) पर अपलोड करने का कष्ट करें।
2. संयुक्त निदेशक (मुख्यालय), स्वास्थ्य भवन, लखनऊ / नोटिस बोर्ड पर चस्पा हेतु।
3. प्रभारी, कम्प्यूटर सेल, कार्यालय, प्रमुख सचिव, चिकित्सा स्वास्थ्य एवं परिवार कल्याण, उ०प्र० शासन।

  
निदेशक (भण्डार)