

# कार्यालय महानिदेशक चिकित्सा एवं स्वास्थ्य सेवायें उ०प्र०

## लखनऊ ।

(राजकीय केन्द्रीय औषधि भण्डार)

पत्र सं०-८फ/८०अनु०-७४१/१९४६

लखनऊ : दिनांक ८ मार्च, २०१६

### “शुद्धि-पत्र”

वर्ष २०१५-१६ में औषधियों के दर अनुबन्ध हेतु महानिदेशालय स्तर से आमंत्रित निविदा सं०-८फ/आर०सी०-७४१/२०६, दिनांक-१४.०१.२०१६ में निम्नतालिकानुसार आंशिक संशोधन किया जाता है :-

विवरण	पूर्व में प्रस्तावित दिनांक व समय	संशोधित दिनांक व समय
निविदा विक्रय का दिनांक	२२ जनवरी, २०१६ से ०८ मार्च २०१६ तक	२२ जनवरी, २०१६ से २८ मार्च, २०१६ तक
नमूना जमा करने की तिथि	०८ मार्च, २०१६ को सायंकाल ०५.०० बजे तक	२८ मार्च, २०१६ को सायंकाल ०५.०० बजे तक
निविदा पड़ने की तिथि	०९ मार्च, २०१६ को अपरान्ह ०२.३० बजे तक	२९ मार्च, २०१६ को अपरान्ह ०२.३० बजे तक
निविदा खोलने की तिथि	०९ मार्च, २०१६ को अपरान्ह ०३.०० बजे तक	२९ मार्च, २०१६ को अपरान्ह ०३.०० बजे तक

शेष नियम व शर्तें यथावत रहेगी।

निदेशक, (भण्डार)

पृष्ठांकन-८फ/८०अनु०-७४१/१९४७-५४

तददिनांक

प्रतिलिपि निम्नलिखित को सूचनार्थ एवं आवश्यक कार्यवाही हेतु प्रेषित :-

- निदेशक, सूचना एवं जन-सम्पर्क विभाग, उ०प्र०, लखनऊ को इस आशय से प्रेषित कि उक्त संशोधित तिथि को तत्काल प्रकाशित करने का कष्ट करें। (आर०ओ०नं०-१०३२६५)
- संयुक्त निदेशक (मुख्यालय), स्वास्थ्य भवन, लखनऊ को इस आशय से प्रेषित कि कृपया उक्त सूचना को महानिदेशालय के सूचना-पट पर चस्पा कराने का कष्ट करें।
- इण्डियन ड्रग मैनुफैक्चरर एसोसियेशन, १०२ पूनम चैम्बर, 'ए' विंग फर्स्ट फ्लोर डा०ए०बी० रोड वारली मुम्बई-४०००१८
- बल्क ड्रग मैनुफैक्चरर एसोसियेशन, सी-२५ इन्डसट्रियल स्टेट सनथ नगर हैदराबाद-५०००१८
- ड्रग नियंत्रक, जगत नारायण रोड उ०प्र० लखनऊ को इस आशय से प्रेषित कि सभी राज्य के ड्रग नियंत्रकों को अपने स्तर से सूचित करने का कष्ट करें।
- ओ०आई०सी०सी० उ०प्र० शासन को इस आशय से प्रेषित कि उपरोक्त निविदा का समय विस्तार विभागीय वेबसाइट पर तत्काल लोड करने का कष्ट करें।
- सम्बन्धित फर्म।
- गार्ड फाइल।

निदेशक (भण्डार)

## Annexure-VII

### ANNUAL TURN OVER STATEMENT

The Annual Turnover of M/s. \_\_\_\_\_ for the past three years are given below and certified that the statement is true and correct.

Sl.No.	Financial Year	Turnover in Crores (Rs)	
1	20XX – 20XX		
2	20XX – 20XX		
3	20XX – 20XX		

Total - Rs. \_\_\_\_\_ Crores

Average turnover per annual - Rs. \_\_\_\_\_ Crores

Date:

Signature of Statutory Auditor/ Chartered Accountant  
(Name in Capital)

Reg. No.-----

### STATEMENT “A”

Name & Specification of the quoted products along with the serial number as per list of tender items.	Number & date of issue of 1 <sup>st</sup> Drug License of the item.	Date of manufacture of 1 <sup>st</sup> batch	Date of 1 <sup>st</sup> sale in the market	Number of batches manufactured in 2007-08, 2008-09
1	2	3	4	5

### STATEMENT “B”

Name & Specification of the quoted products along with the serial number as per list of tender items.	Whether quoted product sugar coated. Firm coated, non sugar coated tabs.	Whether item in strips of aluminum/blister of glass/plastic bottle	Whether UPGS will be embossed on tabs. & printed caps	Whether UPGS NOT FOR SALE will be printed of foil/pack/tab/blister and lable of amps/vials/phial bottle & outer cartons.
1	2	3	4	5

Signature and Seal

## Non Drug Item which are withdrawn.

190	S54	Variable Concentration Mask			. Suitable for oxygen therapy. Swivel connector for patient comfort
191	S55	Aerosol Therapy NEBULIZING KIT	Complete kit consists of : Nebulizer Mask, Nebulizer Chamber and 210 cm Star Lumen Tube.	STERILIZED PKD SINGLE PIECE PACK	. Specially designed nebulizer chamber with nebulization rate of approximately 3 cc in 10 minutes.. Average particle size generation is approximately 2 microns
238	S69	HUB Cutter		STERILIZED PKD SINGLE PIECE PACK	
239	S70	Cord clamp		STERILIZED PKD SINGLE PIECE PACK	
240	S71	cord clamp with cutting facility		STERILIZED PKD SINGLE PIECE PACK	
241	S75q	Xray plate	8x10	PACKETS OF 125	Digital Camera Films (laser type)
242	S75r	Xray plate	10x12	PACKETS OF 125	Digital Camera Films (laser type)
243	S75s	Xray plate	14x17	PACKETS OF 125	Digital Camera Films (laser type)
244	S76	Rota haler device			
245	S77	SWINE FLU VIRAL TRANSPORT MEDIUM (HI VIRAL SWAB WITH 3%HI TRANSPORT MEDIUM			15x120mm polypropylene centrifuge tube, 2ml fill of transport media, nylon  flocked NP swab on a thin flexible plastic shaft that is scored for breaking, includes a zip lock bag,
246	S78	N-95 MASK		STERILIZED PKD SINGLE PIECE PACK	N-95 Face Respirator Mask: Filter efficiency of 95 % or more against particulate aerosols. The mask should be provided with expiration valve. It should be disposable & to be able to fit for wide range of face sizes. It should accompany with certification from NIOSH or any other internationally accepted certification.

247	S79	TRIPPLE LAYER MASK		STERILIZED PKD SINGLE PIECE PACK	Tie on Mask of Non-woven, Hypoallergenic 3 ply construction with filter in between offering >99 percent standard with 4 tie strings.
248	S80	PP KIT LOW WITH SEPARATE HOOD		STERILIZED PKD SINGLE PIECE PACK	Pandemic Flu Kit, LOW GOWN WITH SEPARATE HOOD Including Gloves, Mask, shoe cover 4 Oz. Sanitizing Lotion, Sanitizing Wipes, Tissues And Plastic Resealable Bag.
249	S82	SHOE COVER			
260	S86a	STERILE SURGICAL GLOVES POWDER FREE	FG6	STERILIZED PKD SINGLE PIECE PACK	Box of 100
261	S86b	STERILE SURGICAL GLOVES POWDER FREE	FG6.5	STERILIZED PKD SINGLE PIECE PACK	Box of 100
262	S86c	STERILE SURGICAL GLOVES POWDER FREE	FG7	STERILIZED PKD SINGLE PIECE PACK	Box of 100
263	S86d	STERILE SURGICAL GLOVES POWDER FREE	FG 7.5	STERILIZED PKD SINGLE PIECE PACK	Box of 100
264	S86e	STERILE SURGICAL GLOVES POWDER FREE	FG8	STERILIZED PKD SINGLE PIECE PACK	Box of 100
265	S86f	STERILE SURGICAL GLOVES POWDER FREE	FG8.5	STERILIZED PKD SINGLE PIECE PACK	Box of 100
278	S89	INSULIN NANO PEN		UNIT	
293	S93a	SURGICAL BLADE SS/CS	No.11	STERILIZED PKD SINGLE PIECE PACK	
294	S93b	SURGICAL BLADE SS/CS	No.12	STERILIZED PKD SINGLE PIECE PACK	
295	S93c	SURGICAL BLADE SS/CS	No.15	STERILIZED PKD SINGLE PIECE PACK	
296	S93d	SURGICAL BLADE SS/CS	No.21	STERILIZED PKD SINGLE PIECE	

				<b>PACK</b>	
<b>297</b>	<b>S93e</b>	<b>SURGICAL BLADE SS/CS</b>	<b>No.22</b>	<b>STERILIZED PKD SINGLE PIECE PACK</b>	
<b>298</b>	<b>S93f</b>	<b>SURGICAL BLADE SS/CS</b>	<b>No.23</b>	<b>STERILIZED PKD SINGLE PIECE PACK</b>	
<b>300</b>	<b>S95</b>	<b>DISPOSABLE SURGICAL MASK</b>		<b>STERILIZED PKD SINGLE PIECE PACK</b>	
<b>301</b>	<b>S96</b>	<b>DISPOSABLE SURGICAL CAP</b>		<b>STERILIZED PKD SINGLE PIECE PACK</b>	
<b>302</b>	<b>S97</b>	<b>MACINTOSH SHEET</b>		<b>ROLL</b>	
<b>307</b>	<b>S99a</b>	<b>ECG ROLL</b>	<b>50 cm x 20 Mtr</b>	<b>ROLL</b>	
<b>308</b>	<b>S99b</b>	<b>ECG ROLL</b>	<b>50 cm x 30 Mtr</b>	<b>ROLL</b>	
<b>309</b>	<b>S100</b>	<b>USG ROLL</b>	<b>2-1/16"X250 ft</b>	<b>ROLL</b>	
<b>310</b>	<b>S101</b>	<b>ECG GEL</b>	<b>250ml</b>	<b>TUBE</b>	
<b>337</b>	<b>S106a</b>	<b>STOCKINETTE</b>	<b>5cmX10mm</b>	<b>ROLL</b>	
<b>338</b>	<b>S106b</b>	<b>STOCKINETTE</b>	<b>7.5cmX10mm</b>	<b>ROLL</b>	
<b>339</b>	<b>S106c</b>	<b>STOCKINETTE</b>	<b>10cmX10mm</b>	<b>ROLL</b>	
<b>340</b>	<b>S106d</b>	<b>STOCKINETTE</b>	<b>15cmX10mm</b>	<b>ROLL</b>	
<b>341</b>	<b>S107a</b>	<b>CAST PADDING</b>	<b>2.7mt</b>	<b>ROLL</b>	
<b>342</b>	<b>S107b</b>	<b>CAST PADDING</b>	<b>7.5cmx2.7mt</b>	<b>ROLL</b>	
<b>343</b>	<b>S108b</b>	<b>STEAM AUTOCLAVE LEAD FREE TAPE</b>	<b>3/4"X55mt</b>	<b>ROLL</b>	
<b>344</b>	<b>S108b</b>	<b>STEAM AUTOCLAVE LEAD FREE TAPE</b>	<b>1/2"X30mt</b>	<b>ROLL</b>	
<b>461</b>	<b>S137</b>	<b>hCG Pregnancy Test Device</b>			
<b>464</b>	<b>S140</b>	<b>Sputum container</b>		<b>Cups made of Special Medical Grade Polypropylene, thin plastic, translucent, Diameter - 4cm,Capacity - 30 ml, Srewable Cap should also be made of Special Medical Grade Polypropylene and should be air tight. Leak proof.</b>	
<b>Drug Item which are withdrawn.</b>					
<b>228</b>	<b>S61a</b>	<b>Gauze Than : As per Schedule</b>	<b>60 cm</b>	<b>Each</b>	<b>18 mtr</b>

		<b>F-II of Drugs &amp; Cosmetics Rules, 1945</b>			
<b>229</b>	<b>S61b</b>	<b>Gauze Than : As per Schedule F-II of Drugs &amp; Cosmetics Rules, 1945</b>	<b>45 cm</b>	<b>Each</b>	<b>18 mtr</b>
<b>233</b>	<b>S65a</b>	<b>Bandage Than As per Schedule F-II of Drugs &amp; Cosmetics Rules, 1945</b>	<b>100 cm</b>	<b>Each</b>	<b>20 mtr</b>
<b>234</b>	<b>S65b</b>	<b>Bandage Than As per Schedule F-II of Drugs &amp; Cosmetics Rules, 1945</b>	<b>45 cm</b>	<b>Each</b>	<b>18 mtr</b>
<b>450</b>	<b>S130A</b>	<b>Syphilis Rapid Test Card</b>			
<b>451</b>	<b>S130B</b>	<b>Syphilis Test Strip</b>			
<b>452</b>	<b>S131A</b>	<b>Dengue NS -1 Ag +IgG Ab Test Card</b>			
<b>453</b>	<b>S131B</b>	<b>Dengue NS -1 Ag +IgM Ab Test Card</b>			
<b>456</b>	<b>S133A</b>	<b>Malaria Ag Pf</b>			
<b>457</b>	<b>S133B</b>	<b>Malaria Ag Pv</b>			
<b>Item below require latest BIS Specification Certificate.</b>					
<b>236</b>	<b>S67</b>	<b>Insuline Syringe Fixed Needle</b>	<b>1 ml, 40 U</b>	<b>STERILIZED PKD SINGLE PIECE PACK</b>	<b>Each syringe should separately HDPE card board packing and be Gamma sterilized E.T.O. the washer of the syringe should be non toxic medical grade. Needle should be properly protected. The product and package should be designed as an integrated unit. The package should be sealed with a good heat sealing machine so that the seal with is not less than 2 to 3 mm the body of the packing material as well as the seal should not contain pinholes.</b>
<b>252</b>	<b>S85a</b>	<b>DISPOSABLE NEEDLE</b>	<b>FG 16</b>	<b>STERILIZED PKD SINGLE PIECE PACK</b>	
<b>253</b>	<b>S85b</b>	<b>DISPOSABLE NEEDLE</b>	<b>FG 18</b>	<b>STERILIZED PKD SINGLE PIECE PACK</b>	

254	S85c	DISPOSABLE NEEDLE	FG 20	STERILIZED PKD SINGLE PIECE PACK	
255	S85d	DISPOSABLE NEEDLE	FG 22	STERILIZED PKD SINGLE PIECE PACK	
256	S85e	DISPOSABLE NEEDLE	FG 23	STERILIZED PKD SINGLE PIECE PACK	
257	S85f	DISPOSABLE NEEDLE	FG 24	STERILIZED PKD SINGLE PIECE PACK	
258	S85g	DISPOSABLE NEEDLE	FG 25	STERILIZED PKD SINGLE PIECE PACK	
259	S85h	DISPOSABLE NEEDLE	FG 26	STERILIZED PKD SINGLE PIECE PACK	
272	88a	DISPOSABLE SYRINGE WITHOUT NEEDLE	10 ml	STERILIZED PKD SINGLE PIECE PACK	Each syringe should separately HDPE card board packing and be Gamma sterilized E.T.O. the washer of the syringe should be non toxic medical grade. The product and package should be designed as an integrated unit. The package should be sealed with a good heat sealing machine so that the seal with is not less than 2 to 3 mm the body of the packing material as well as the seal should not contain pinholes.
273	88b	DISPOSABLE SYRINGE WITHOUT NEEDLE	2 ml	STERILIZED PKD SINGLE PIECE PACK	Each syringe should separately HDPE card board packing and be Gamma sterilized E.T.O. the washer of the syringe should be non toxic medical grade. The product and package should be designed as an integrated unit. The package should be sealed with a good heat sealing machine so that the seal with is not less than 2 to 3 mm the body of the packing material as well as the seal should not contain pinholes.
274	88c	DISPOSABLE SYRINGE WITHOUT NEEDLE	3 ml	STERILIZED PKD SINGLE PIECE PACK	Each syringe should separately HDPE card board packing and be Gamma sterilized E.T.O. the washer of the syringe should be non toxic medical grade. The product and

					<p>package should be designed as an integrated unit. The package should be sealed with a good heat sealing machine so that the seal with is not less than 2 to 3 mm the body of the packing material as well as the seal should not contain pinholes.</p>
275	S88d	DISPOSABLE SYRINGE WITHOUT NEEDLE	20ml	STERILIZED PKD SINGLE PIECE PACK	<p>Each syringe should separately HDPE card board packing and be Gamma sterilized E.T.O. the washer of the syringe should be non toxic medical grade. The product and package should be designed as an integrated unit. The package should be sealed with a good heat sealing machine so that the seal with is not less than 2 to 3 mm the body of the packing material as well as the seal should not contain pinholes.</p>
276	S88e	DISPOSABLE SYRINGE WITHOUT NEEDLE	5 ml	STERILIZED PKD SINGLE PIECE PACK	<p>Each syringe should separately HDPE card board packing and be Gamma sterilized E.T.O. the washer of the syringe should be non toxic medical grade. The product and package should be designed as an integrated unit. The package should be sealed with a good heat sealing machine so that the seal with is not less than 2 to 3 mm the body of the packing material as well as the seal should not contain pinholes.</p>
277	S88f	DISPOSABLE SYRINGE WITHOUT NEEDLE	50 ml	STERILIZED PKD SINGLE PIECE PACK	<p>Each syringe should separately HDPE card board packing and be Gamma sterilized E.T.O. the washer of the syringe should be non toxic medical grade. The product and package should be designed as an integrated unit. The package should be sealed with a good heat sealing machine so that the seal with is not less than 2 to 3 mm the body of the packing material as well as the seal should not contain pinholes.</p>



279	S90a	INSULIN NEEDLE	26G	STERILIZED PKD SINGLE PIECE PACK	4 mm
280	S90b	INSULIN NEEDLE	28G FOR PEN	STERILIZED PKD SINGLE PIECE PACK	4 mm
281	S90c	INSULIN NEEDLE	30G FOR PEN	STERILIZED PKD SINGLE PIECE PACK	4 mm
282	S90d	INSULIN NEEDLE	32G FOR PEN	STERILIZED PKD SINGLE PIECE PACK	4 mm
289	S92a	SPINAL NEEDLE	FG22	STERILIZED PKD SINGLE PIECE PACK	
290	S92b	SPINAL NEEDLE	FG23	STERILIZED PKD SINGLE PIECE PACK	
291	S92c	SPINAL NEEDLE	FG24	STERILIZED PKD SINGLE PIECE PACK	
292	S92d	SPINAL NEEDLE	FG25	STERILIZED PKD SINGLE PIECE PACK	
348	S110a	DISPOSABLE SYRINGE ALONG WITH NEEDLE	1 ml	STERILIZED PKD SINGLE PIECE PACK	Each syringe should separately HDPE card board packing and be Gamma sterilized E.T.O. the washer of the syringe should be non toxic medical grade. Needle should be properly protected. The product and package should be designed as an integrated unit. The package should be sealed with a good heat sealing machine so that the seal with is not less than 2 to 3 mm the body of the packing material as well as the seal should not contain pinholes.
349	S110b	DISPOSABLE SYRINGE ALONG WITH NEEDLE	10 ml	STERILIZED PKD SINGLE PIECE PACK	Each syringe should separately HDPE card board packing and be Gamma sterilized E.T.O. the washer of the syringe should be non toxic medical grade. Needle should be properly protected. The product and package should be designed as an integrated unit. The package should be sealed with a good heat sealing machine so that the seal with is not less than 2 to 3 mm the body of the packing material as

					well as the seal should not contain pinholes.
350	S110c	DISPOSABLE SYRINGE ALONG WITH NEEDLE	2 ml	STERILIZED PKD SINGLE PIECE PACK	Each syringe should separately HDPE card board packing and be Gamma sterilized E.T.O. the washer of the syringe should be non toxic medical grade. Needle should be properly protected. The product and package should be designed as an integrated unit. The package should be sealed with a good heat sealing machine so that the seal with is not less than 2 to 3 mm the body of the packing material as well as the seal should not contain pinholes.
351	S110d	DISPOSABLE SYRINGE ALONG WITH NEEDLE	3 ml	STERILIZED PKD SINGLE PIECE PACK	Each syringe should separately HDPE card board packing and be Gamma sterilized E.T.O. the washer of the syringe should be non toxic medical grade. Needle should be properly protected. The product and package should be designed as an integrated unit. The package should be sealed with a good heat sealing machine so that the seal with is not less than 2 to 3 mm the body of the packing material as well as the seal should not contain pinholes.
352	S110e	DISPOSABLE SYRINGE ALONG WITH NEEDLE	20ml	STERILIZED PKD SINGLE PIECE PACK	Each syringe should separately HDPE card board packing and be Gamma sterilized E.T.O. the washer of the syringe should be non toxic medical grade. Needle should be properly protected. The product and package should be designed as an integrated unit. The package should be sealed with a good heat sealing machine so that the seal with is not less than 2 to 3 mm the body of the packing material as well as the seal should not contain pinholes.
353	S110f	DISPOSABLE SYRINGE ALONG WITH NEEDLE	5 ml	STERILIZED PKD SINGLE PIECE PACK	Each syringe should separately HDPE card board packing and be Gamma sterilized E.T.O. the washer of the syringe

					<p>should be non toxic medical grade. Needle should be properly protected. The product and package should be designed as an integrated unit. The package should be sealed with a good heat sealing machine so that the seal with is not less than 2 to 3 mm the body of the packing material as well as the seal should not contain pinholes.</p>
354	S110g	DISPOSABLE SYRINGE ALONG WITH NEEDLE	50 ml	STERILIZED PKD SINGLE PIECE PACK	<p>Each syringe should separately HDPE card board packing and be Gamma sterilized E.T.O. the washer of the syringe should be non toxic medical grade. Needle should be properly protected. The product and package should be designed as an integrated unit. The package should be sealed with a good heat sealing machine so that the seal with is not less than 2 to 3 mm the body of the packing material as well as the seal should not contain pinholes.</p>