

कार्यालय महानिदेशक, चिकित्सा एवं स्वास्थ्य सेवायें, उ०प्र० लखनऊ।

(सी०एम०एस०डी०, अनुभाग-८)

Phone no: - 0522-2625809

E-mail id: - cmsd.dghealth@gmail.com

संख्या:-८फ/क्यू०सी०-७९१/

लखनऊ: दिनांक: १९ जुलाई, २०१६

शुद्धि पत्र

महानिदेशालय के पत्र सं०-८फ/क्यू०सी०-७९१/१७२९ दिनांक २४.०६.२०१६ द्वारा मात्रा अनुबंध के अंतर्गत जनपद कानपुर नगर, बाराबंकी, हरदोई, आजमगढ़, बुलंदशहर, फिरोजाबाद, उन्नाव, सुल्तानपुर, गाजियाबाद, डा० राम मनोहर लोहिया संयुक्त चिकित्सालय लखनऊ, जौनपुर, कन्नौज, सोनभद्र, बांदा, बलिया, पं० दीनदयाल उपाध्याय चिकित्सालय वाराणसी एवं एस०बी०डी० चिकित्सालय सहारनपुर (राज्य बजट से) तथा उरई, ललितपुर, फैजाबाद, फतेहपुर, इटावा, बस्ती (एन०एच०एम० बजट से) में ट्रामा सेन्टर हेतु उपकरणों की आपूर्ति/स्थापना कराये जाने हेतु निविदा आमंत्रित की गई थी।

उक्त निविदा में मॉनीटर लार्ज स्क्रीन, विड ई०सी०जी०, एस०पी०ओ०टू०, एन०आई०बी०पी० एण्ड ई०टी०सी०ओ०टू० के मानकों को निम्नानुसार संशोधित किया जाता है:-

MONITOR LARGE SCREEN

TECHNICAL SPECIFICATIONS

1. Should be suitable for adult, Pediatric, neonatal patients monitoring.
2. Monitor should have facility to display 5 Para like ECG, RR, HR, SPO2 (Nellcor), NIBP Dual and as standard Parameters with built in rechargeable battery, capric probe backup of at least 2 hours & recorder for continuous operation.
3. Display TFT Display of size not less than 15". Should display at least 5 waveforms of selected parameters simultaneously.
4. It should be able to analyze arrhythmia, ST-segments, ST analysis, HRV, NIBP analysis etc.
5. Monitor should have large font facility.
6. ETCO2 Monitoring should be provided with each monitor.
7. Should be able to store & display up to 72 hrs. Of graphical trends of all parameters and up to 40 events with waveform.
8. The monitor should be European CE from notified body and USFDA approved. Attach copy of certificate with bid.
9. Should have inbuilt three channel recorder as standard.
10. The monitor should have built in facility connectivity to Central Station through, Wired/Wireless OR Ethernet Card.
11. Provision for Universal Serial Bus port for software upgrade.
12. Should have battery backup of 120 min.
13. Warranty for whole monitor should be for 3 yrs and CMC for 5 Years.
14. Should have Service Centre in State & the bidder should supply 100 electrode sets per year.
15. Minimum real-time 5 leads ECG.
16. Monitor should be ready to upgradable feature with Cardiac Output, Multi gas Monitoring.
17. Following Standard Accessories to be supplied
 - a) 5 lead ECG Cable - 06 Nos.
 - b) NIBP Cuff (Adult, Pediatric & Neonatal) & tubing - 02 each
 - c) Nellcor SpO2 (Adult & Y probe) with extension cable - 02 each
 - d) Power Cord with earthing cable.
18. Should provide stabilized electricity to all monitors.
19. Supplier to perform installation, safety and operation checks before handover.
20. Local clinical staff to affirm completion of installation.
21. Training of users in operation and basic maintenance shall be provided.

22. List to be provided of important spares and accessories, with their part numbers and cost.

23. Certificate of calibration and inspection to be provided.

24. Documentation:

Should provide 1 set (Hard Copy) to user and 1 Set (Soft copy) by email to cmsd.dghealth@gmail.com with cc to bme.uphssp@gmail.com for centralized record of:

- User, technical, maintenance and service manuals to be supplied along with machine diagrams.
- List of equipment and procedures required for local calibration and routine maintenance.
- Certificate of calibration and inspection.

समय सारणी निम्नवत् संशोधित की जाती है :-

<u>क्यू0सी0-791</u>	
<u>विवरण</u>	<u>दिनांक व समय</u>
निविदा विक्रय की तिथि	01 जुलाई, 2016 से 25 जुलाई, 2016
निविदा पड़ने की तिथि	26 जुलाई, 2016 को अपराह्न 2.30 बजे तक
निविदा खुलने की तिथि	26 जुलाई, 2016 को अपराह्न 3.00 बजे

निविदा की शेष नियम व शर्तें यथावत् रहेंगी।

निदेशक (भण्डार)

संख्या:-8फ/क्यू0सी0-791/2153-55 तददिनांक।

प्रतिलिपि निम्नलिखित को सूचनार्थ एवं आवश्यक कार्यवाही हेतु प्रेषित :-

- इं0 अवधेश रावत, एन.आई.सी. उ0प्र0 लखनऊ को इस आशय से प्रेषित कि निविदा की संशोधित सूचना को वेबसाइट :- www.dghealth.up.nic.in एवं www.uphealth.up.nic.in पर अपलोड करने का कष्ट करें।
- संयुक्त निदेशक (मुख्यालय), स्वास्थ्य भवन, लखनऊ / नोटिस बोर्ड पर चस्पा हेतु।
- प्रभारी, कम्प्यूटर सेल, कार्यालय, प्रमुख सचिव, चिकित्सा स्वास्थ्य एवं परिवार कल्याण, उ0प्र0 शासन।

निदेशक (भण्डार)