

कार्यालय महानिदेशक, चिकित्सा एवं स्वास्थ्य सेवार्ये, 30प्र0 लखनऊ।

(राजकीय केन्द्रीय औषधि भण्डार, अनुभाग-8)

Phone no: - 0522-2625809

E-mail id: - cmsd.dghealth@gmail.com

संख्या:-8फ/क्यू0सी0-756/

लखनऊ: दिनांक: 6 अप्रैल, 2016

:: शुद्धि पत्र ::

महानिदेशालय पत्र संख्या-8फ/क्यू0सी0-756/2026, दिनांक 11.03.2016 द्वारा मात्रा अनुबंध के अंतर्गत उपकरणों की निविदा निर्गत की गयी थी।

उपरोक्त निविदा में निम्नलिखित उपकरणों के मानकों में भारत सरकार की गाइड लाइन के अनुसार निम्नवत् संशोधित किया जाता है :-

Sl. No.	Item Name	Specification in Tender Documents	Revised Specifications
1	3- Ventilator with high end Compressor Or Internal (Turbine Base) Compressor	<p>Ventilator with high end Compressor Or Internal (Turbine Base) Compressor.</p> <p>4- Should be inbuilt compressor/turbine technology and Battery.</p> <p>6- Modes (Breath Delivery)</p> <ul style="list-style-type: none"> > V-CMV (V=Volume) > V-SIMV > Spontaneous > P-CMV (P-Pressure) > P-SIMV > BIPAP/PEEP > APV-CMV > APV-SIMV > APRV > PRVC > NIV <p>8- PEEP/CPAP: UP to 30 cm H2O.</p> <p>9a- Pressure Triggering 1-20 cm H2O.</p> <p>16a-Pipeline Pressure failure.</p> <p>19- The quoted model should have EUROPEAN CE Certificate or USFDA Approved</p> <p>20- Should have installation in any Govt. Hospital Ambulance of U.P.</p>	<p>Ventilator with high end Compressor of same make.</p> <p>4-Should be inbuilt compressor/turbine technology and Battery Backup of minimum 01 hour.</p> <p>6-Modes (Breath Delivery)</p> <ul style="list-style-type: none"> > V-CMV (V=Volume) > V-SIMV > Spontaneous > P-CMV (P-Pressure) > P-SIMV > BIPAP/PEEP > APRV > NIV <p>8- PEEP/CPAP: UP to 60 cm H2O.</p> <p>9a- Pressure Triggering (-20) to 0 cm H2O.</p> <p>16a-Pipeline Pressure failure of O2 and/or air, if applicable.</p> <p>19- The quoted model should be USFDA/EUROPEAN CE/BIS Approved</p> <p>20- NIL</p>
2	Anaesthesia machine with monitor parameters: Agent monitoring NIBP, SPO ₂ , ET CO ₂ , ECG Temp., IBP	<p>Specifications of Multi parameter monitor for ANAESTHESIA MACHINE erroneously not attached in tender documents.</p>	<p>1- The monitor should be modular type and measure Basic Parameters like 3/5 Lead ECG, Resp, Dual Temp, SpO₂ for low perfusion, NIBP.</p> <p>2- Should be capable to be up-</p>

			<p>gradable to Et CO₂; CVP & IBP.</p> <p>3- The monitor should have highly visible, bright 10" or more color TFT touch screen display for easy viewing from distance.</p> <p>4- The monitor should display at least 8 wave form traces on a single screen.</p> <p>5- The monitor should have flexible display mode for various monitoring requirement. Bigger font of numeric should also be available.</p> <p>6- There should be provision for using LAN card & fixed compact flash memory card with power off storage function.</p> <p>7- There should be alarm limit setting for every parameter.</p> <p>8- It should have emergency nurse call alarm function and Defib-Synch function.</p> <p>9- It should have standby mode for temporary leaving of patient.</p> <p>10- There should be complete ST segment, Arrhythmia analysis with Pace detection facility.</p> <p>11- The monitor should have oxy CRG screen.</p> <p>12- It should come with rechargeable battery with 2 hours of battery backup.</p> <p>13- It should have facility of Multi Gas module (Anaesthetic Gases) to be quoted as option (Separately) for use with Anaesthesia Machine.</p> <p>14- The monitor system should be US FDA/EUROPEAN CE/BIS Approved.</p>
3	Fully automatic biochemistry analyzer	USFDA/CE EUROPEAN	USFDA/EUROPEAN CE/BIS Approved
4	500 mA X-ray machine	The offered model must be approved by AERB and CE EUROPEAN CERTIFIED	The offered model must be approved by AERB and USFDA/EUROPEAN CE/BIS Approved
5	Digital X-ray machine	106- Offered system must be USFDA and CE EUROPEAN Approved	106- Offered system must be USFDA/EUROPEAN CE/BIS Approved
6	Colour Doppler Ultrasound Machine	28- The quoted model should be CE Marked and USFDA Approved	28- The quoted model should be USFDA/EUROPEAN CE/BIS Approved
7	O.T. Table- 3 Segments Translucent top with orthopaedic attachment	Quality certificate like CE European, ISO 13485, OHSAS 18001 to be enclosed	Quality certificate like CE European/BIS, with ISO 13485, OHSAS 18001 to be enclosed

8	O.T. Ceiling Light- shadow less	7- Certificate- CE European and ISO 9001, 14001, 13485, OHSAS 18001	7- Certificate- USFDA/EUROPEAN CE/BIS Approved.
9	USG Machine	16- Should be CE European and USFDA Approved.	16- Should be USFDA/EUROPEAN CE/BIS Approved.
10	Orthopaedic Power drill & power saw	8- Should be CE European/USFDA Approved.	8- Should be USFDA/EUROPEAN CE/BIS Approved.
11	Pneumatic Tourniquet	15- Should be CE European/USFDA Approved.	15- Should be USFDA/EUROPEAN CE/BIS Approved.
12	Transport Ventilator	6a- Product should be USFDA Approved b- Comprehensive warranty for three years and provision of AMC for next three years	6a- Product should be USFDA/EUROPEAN CE/BIS Approved. b- Comprehensive warranty for three years and provision of AMC for next five years.

उपरोक्त के अतिरिक्त कम संख्या-1, 2, 3, 4, 5, 6, 9 एवं 12 में उल्लिखित उपकरणों के मानकों के साथ निम्नलिखित मानकों को भी पढ़ा जाय :-

- 1- Manufacturer and supplier should have ISO 13485 certification for quality standard.
- 2- Shall meet Internationally recognized standard for electromagnetic compatibility (EMI/EMC) for electro medical equipment: 61326-1
- 3- Certified to be compliant with IEC61010-1-3, IEC61010-1-2, IEC61010-2-54, IEC61010-1-6 and IEC62304.

उक्त निविदा की समय सारणी को निम्नवत् संशोधित किया जाता है :-

क्यू0सी0-756	
विवरण	दिनांक व समय
निविदा विक्रय की तिथि	17 मार्च, 2016 से 18 अप्रैल, 2016
निविदा पढ़ने की तिथि	19 अप्रैल, 2016 को अपराह्न 2.30 बजे तक
निविदा खुलने की तिथि	19 अप्रैल, 2016 को अपराह्न 3.00 बजे

शेष नियम व शर्तें पूर्ववत् रहेंगी।

(सुरेश चन्द्र)
निदेशक (भण्डार)

संख्या:-8फ/क्यू0सी0-756/ 35-37

तददिनांक।

प्रतिलिपि निम्नलिखित को सूचना संलग्नक सहित सूचनार्थ एवं आवश्यक कार्यवाही हेतु प्रेषित :-

1. इं0 अवधेश रावत, एन.आई.सी. उ0प्र0 लखनऊ को इस आशय से प्रेषित कि निविदा की संशोधित सूचना को वेबसाइट :-www.upgov.nic.in and www.uphealth.up.nic.in पर अपलोड करने का कष्ट करें।
2. संयुक्त निदेशक (मुख्यालय), स्वास्थ्य भवन, लखनऊ /नोटिस बोर्ड पर चरपा हेतु।
3. प्रभारी, कम्प्यूटर सेल, कार्यालय, प्रमुख सचिव, चिकित्सा स्वास्थ्य एवं परिवार कल्याण, उ0प्र0 शासन।

(सुरेश चन्द्र)
निदेशक (भण्डार)