

कार्यालय महानिदेशक, चिकित्सा एवं स्वास्थ्य सेवायें, उ०प्र०, लखनऊ।
(सी०एम०एस०डी०, अनुभाग-८)

Phone no: - 0522-2625809

E-mail id: - cmsd.dghealth@gmail.com

संख्या:-८फ/क्यू०सी०-९३४/

लखनऊ: दिनांक: ०५ जून, 2018

संशोधित निविदा सूचना-२

(केवल ई-टेंडरिंग वेबसाइट www.etender.up.nic.in के माध्यम से)

महानिदेशालय पत्र संख्या-८फ/क्यू०सी०-९३४/७८६, दिनांक १८ मई, २०१८ द्वारा अनुबंध की तिथि से ०१ वर्ष तक मात्रा अनुबंध के अन्तर्गत राष्ट्रीय स्वास्थ्य मिशन द्वारा वित्त पोषित नेशनल वेक्टर बार्न डिजीज कन्ट्रोल प्रोग्राम के अन्तर्गत ०६ जनपदों आजमगढ़, गोण्डा, रायबरेली, हरदोई, सीतापुर एवं मऊ के जिला/जिला संयुक्त चिकित्सालयों में १० शैय्यायुक्त एवं ०५ वेन्टीलेटर युक्त पीडियाट्रिक आई०सी०यू० की स्थापना कराये जाने हेतु उपकरणों के क्रय, आपूर्ति/स्थापना कराये जाने हेतु ख्याति प्राप्त निर्माता/निर्माता फर्मों के अधिकृत डीलरों/निर्माता फर्मों के अधिकृत एजेंटों से निविदाएँ ई-टेंडर के माध्यम से आमंत्रित की गयी थी।

उक्त निविदा के अन्तर्गत निम्नवत् संशोधन किया जाता है :-

Technical Specification as per Tender	Technical Specification after amendment
	(1) ICU BEDS
1. It should have minimum seven functionality - Back Rest/Knee Rest /Trendelenburg / Reverse Trendelenburg /High-Low / Back Rest-Knee Rest Simulataneoulsy/ Cardiac Chair Position/Automatic CPR and special Neuro Position, operated by separate actuators by hand held remote control system to adjust the bed into a comfortable position.	1. It should have minimum seven functionality - Back Rest/Knee Rest /Trendelenburg / Reverse Trendelenburg /High-Low / Back Rest-Knee Rest Simulataneoulsy/ Cardiac Chair Position/Automatic CPR and Manual CPR, operated by separate actuators by hand held remote control system to adjust the bed into a comfortable position.
3. It should have following technical features :- Standard Length of approx: 2200 – 2350 mm (215-220 cm) , Standard Width of approx: 1025 -1035 mm, Standard Top height with the available 120-130 mm castors approx: 750-765 mm, Standard Low height with the available 120-130 mm castors approx: 315 mm- 325 mm, with a capacity of Safe working load/Maximum Patient Weight – 150/ 250 kg, Standard Side Railing : high duty ABS , Standard Railing area covering : 90%-95 of the bed length for achieving total safety to the Patient, I.V Saline Stand Provision : 04 locations and should have facility to insert into the sleeves on either side of the bed, Mattress guard : metal with a property of non corrosive for achieving proper terminal disinfection thru low level disinfection solution Base of ICU Bed : fully cover to allow safe access for terminal cleaning & fumigation of ICU and decontamination.	3. It should have following technical features :- Standard Length of approx: 2200 – 2350 mm (215-220 cm) , Standard Width of approx: 975-1025 mm, Standard Top height with the available 120-130 mm castors approx: 750-765 mm, Standard Low height with the available 120-130 mm castors approx: 400 mm- 500 mm, with a capacity of Safe working load/Maximum Patient Weight – 150/ 250 kg, Standard Side Railing : high duty ABS , Standard Railing area covering : 90%-95 of the bed length for achieving total safety to the Patient, I.V Saline Stand Provision : 04 locations and should have facility to insert into the sleeves on either side of the bed, Mattress guard : metal with a property of non corrosive for achieving proper terminal disinfection thru low level disinfection solution Base of ICU Bed : fully cover to allow safe access for terminal cleaning & fumigation of ICU and decontamination.
(2) BED SIDE MONITOR WITH CENTRAL MONITORING STATION	
1. F. Optional item to be quoted: invasive blood pressure-monitoring module complete with reusable transducer, CVP-monitoring.	1. F. Invasive blood pressure-monitoring module complete with reusable transducer, CVP-monitoring module, EtCO2 Monitoring module.
1. D. Capable of saving data for min 24 hrs.	1. D. Capable of saving data for min 72 hrs.
4.1 Accessories: 2 pairs, 12 lead ECG cable. 2 packs of 100 disposable ECG connection electrodes. Two sets of	4.1 Accessories: 2 pairs, 3 lead ECG cable. 2 packs of 100 disposable ECG connection electrodes. Ten SpO2 probes of pediatric & neonatal

reusable SpO2 probes including adult, pediatric & neonatal probes. Two sets of NIBP cuffs of each size. Two external skin temperature probes, CVP-Cable and transducer.	probes. Ten sets of NIBP cuffs of each size. Two external skin temperature probes, CVP-Cable and transducer.
(3) DEFIBRILLATOR	
4.4 100 gel sheet or pads for monitoring and defibrillation.	4.4 50 nos. of disposable pads for AED facility for each defibrillator.
(4) CENTRAL GAS PIPELINE (Oxygen, compressed air, vaccume) (2 points. for O2 and 1 for vaccume and compressed air)	
6.0 Terminal units (Gas Outlets) with adaptors for O2 C. Air, Vacuum (Indigenous)- 140 each Alternately: Terminal Units (Gas outlets) with adaptors for O2, C. Air.	6.0 Terminal units (Gas Outlets) with adaptors for O2 C. Air, Vacuum (Indigenous)- 48 each Alternately: Terminal Units (Gas outlets) with adaptors for O2, C. Air.

निविदा की शेष नियम व शर्तें यथावत् रहेंगी।

निदेशक (भण्डार)

संख्या:-8फ/क्यू0सी0-934/१५३-५५ तद्दिनांक:

प्रतिलिपि निम्नलिखित को सूचनार्थ एवं आवश्यक कार्यवाही हेतु प्रेषित :-

1. संयुक्त निदेशक (मुख्यालय), स्वास्थ्य भवन, लखनऊ / नोटिस बोर्ड पर चस्पा हेतु।
2. प्रभारी कम्प्यूटर सेल, कार्यालय प्रमुख सचिव, चिकित्सा स्वास्थ्य एवं परिवार कल्याण, उत्तर प्रदेश शासन।

निदेशक (भण्डार)