

कार्यालय महानिदेशक, चिकित्सा एवं स्वास्थ्य सेवार्ये, 30प्र0 लखनऊ।

(राजकीय केन्द्रीय औषधि भण्डार, अनुभाग-8)

Phone no: - 0522-2625809

E-mail id: - cmsd.dghealth@gmail.com

संख्या:-8फ/क्यू0सी0-756/

लखनऊ: दिनांक: 12 अप्रैल, 2016

:: शुद्धि पत्र ::

महानिदेशालय पत्र संख्या-8फ/क्यू0सी0-756/2026, दिनांक 11.03.2016 द्वारा मात्रा अनुबंध के अंतर्गत उपकरणों की निविदा निर्गत की गयी थी।

उपरोक्त निविदा में निम्नलिखित उपकरण के मानकों में भारत सरकार की गाईड लाइन के अनुसार निम्नवत् संशोधित किया जाता है :-

Sl. No.	Item Name	Specification in Tender Documents	Revised Specifications
1	3- Ventilator with high end Compressor Or Internal (Turbine Base) Compressor	Ventilator with high end Compressor Or Internal (Turbine Base) Compressor. 4- Should be inbuilt compressor/turbine technology and Battery. 6- Modes (Breath Delivery) ➤ V-CMV (V=Volume) ➤ V-SIMV ➤ Spontaneous ➤ P-CMV (P-Pressure) ➤ P-SIMV ➤ BIPAP/PEEP ➤ APV-CMV ➤ APV-SIMV ➤ APRV ➤ PRVC ➤ NIV 8- PEEP/CPAP: UP to 30 cm H2O. 9a- Pressure Triggering 1-20 cm H2O. 16a-Pipeline Pressure failure. 19- The quoted model should have EUROPEAN CE Certificate or USFDA Approved 20- Should have installation in any Govt. Hospital Ambulance of U.P.	Ventilator with high end Compressor of same make. 4-Should be inbuilt compressor/turbine technology and Battery Backup of minimum 01 hour. 6-Modes (Breath Delivery) ➤ V-CMV (V=Volume) ➤ V-SIMV ➤ Spontaneous ➤ P-CMV (P-Pressure) ➤ P-SIMV ➤ BIPAP/PEEP ➤ APRV ➤ NIV 8- PEEP/CPAP: UP to 60 cm H2O. 9a- Pressure Triggering (-20) to 0 cm H2O. 16a-Pipeline Pressure failure of O2 and/or air, if applicable. 19- The quoted model should be USFDA/EUROPEAN CE/BIS Approved 20- NIL

उपरोक्त उपकरण के मानक में पुनः निम्नवत् संशोधन किया जाता है :-

Sl. No.	Item Name	Specification in Tender Documents	Revised Specifications
1	3- Ventilator with high end Compressor Or Internal (Turbine Base) Compressor	Ventilator with high end Compressor Or Internal (Turbine Base) Compressor.	1- Ventilator with high end Compressor of same make. 2- USFDA/EUROPEAN CE/BIS

शेष नियम व शर्तें पूर्ववत् रहेंगी।

(सुरेश चन्द्र)
निदेशक (भण्डार)

संख्या:-8फ/क्यू0सी0-756/

83-85

तददिनांक।

प्रतिलिपि निम्नलिखित को सूचना संलग्नक सहित सूचनार्थ एवं आवश्यक कार्यवाही हेतु प्रेषित :-

- इं0 अवधेश रावत, एन.आई.सी. 30प्र0 लखनऊ को इस आशय से प्रेषित कि निविदा की संशोधित सूचना को वेबसाइट :-www.upgov.nic.in and www.uphealth.up.nic.in पर अपलोड करने का कष्ट करें।
- संयुक्त निदेशक (मुख्यालय), स्वास्थ्य भवन, लखनऊ /नोटिस बोर्ड पर चस्पा हेतु।
- प्रभासी, कम्प्यूटर सेल, कार्यालय, प्रमुख सचिव, चिकित्सा स्वास्थ्य एवं परिवार कल्याण, 30प्र0 शासन।

(सुरेश चन्द्र)
निदेशक (भण्डार)