

कायालय महानदशक, चिकित्सा एव स्वास्थ्य सवाय, उ०प्र० लखनऊ ।

(सी०एम०एस०डी०, अनुभाग-८)

Phone no: - 0522-2625809

E-mail id: - cmsd.dghealth@gmail.com

संख्या:-८फ/क्यू०सी०-८७५/

लखनऊ: दिनांक: २० मार्च, २०१७

संशोधित निविदा सूचना

महानिदेशालय पत्र संख्या-८फ/क्यू०सी०-८७५/१०६५२, दिनांक ०८.०३.२०१७ द्वारा प्रदेश में रेडियोलोजी उपकरणों की निविदायें आमंत्रित की गयी थी।

उक्त निविदा के मानकों में निम्नवत् संशोधन किये जाते हैं :-

Page No.	Existing Heading / Specification	Revised (To be Read as)
1, 2	X-ray Machine (100MM) The Developing Tank, safelight, X-ray Dark Room, Cassettes X-ray, Intensifying Screen, Chest Stand X-ray (New 42 CHC +6 New CHC) & Ultrasound Machine (New 42 CHC +6 New CHC)	X-ray Machine (100mA) with Automatic film Processor for (New 42 CHC +6 New CHC) & Ultrasound Machine for (New 42 CHC +6 New CHC)
4	2. एक्स-रे मशीन (100 एमएम) व डेवलपिंग टंक, सेफ लाइट, एक्स-रे डार्क रूम, कैसेट्स एक्स-रे, इन्टेन्सिफाइंग स्क्रीन, चैस्ट स्टैंड एक्स-रे (नवनिर्मित 42 नग + 06 नग नये सामु० रवा० केन्द्रों)	2. एक्स-रे मशीन (100 एमएम) विद आटोमेटिक फिल्म प्रोसेसर (नवनिर्मित 42 नग + 06 नग नये सामु० रवा० केन्द्रों) के लिए
5	2. X-ray Machine (100MM) The Developing Tank, safelight, X-ray Dark Room, Cassettes X-ray, Intensifying Screen, Chest Stand X-ray (New 42 CHC +6 New CHC)	2. X-ray Machine (100mA) with Automatic film Processor for (New 42 CHC +6 New CHC)
34	2) System should be offered with following Broad Band width transducers: d) Multi Frequency Convex Array Transducer (frequency range of 2 to 8 Mhz) for Abdominal, Ob/Gyn applications as standard. e) Multi Frequency Linear Array Transducer (frequency range of 5 to 12 Mhz) for vascular & small parts applications (Should be quoted optionally). f) Multi frequency Endocavity Transducer (frequency range of 4 to 9 Mhz) for TV/TR applications, (Should be quoted optionally). 13) System should have a high resolution articulating non-interlaced flicker free antiglare flat panel 15 inches Color LED Monitor.	1)System should be offered with following Broad Band width transducers: a) Multi Frequency Convex Array Transducer (frequency range of 2 to 5 Mhz) for Abdominal, Ob/Gyn applications as standard. b) Multi Frequency Linear Array Transducer (frequency range of 4 to 10 Mhz) for vascular & small parts applications (Should be quoted optionally). c) Multi frequency Endocavity Transducer (frequency range of 4 to 8 Mhz) for TV/TR applications, (Should be quoted optionally). 12) System should have a medical grade high resolution articulating non-interlaced flicker free antiglare flat panel minimum 17 inches or more Color LCD/LED Monitor.

निविदा की शेष नियम व शर्तें यथावत् रहेगी।

अपर निदेशक (भण्डार)

संख्या:-८फ/क्यू०सी०-८७५/१०७९५-९६

तद्दिनांक।

प्रतिलिपि निम्नलिखित को सूचनार्थ एवं आवश्यक कार्यवाही हेतु प्रेषित :-

- इं० अवधेश रावत, एन.आई.सी. उ०प्र० लखनऊ को इस आशय से प्रेषित कि निविदा की संशोधित सूचना को वेबसाइट :- www.dghealth.up.nic.in एवं www.uphealth.up.nic.in पर अपलोड करने का कष्ट करें।
- संगुक्त निदेशक (मुख्यालय), स्वास्थ्य भवन, लखनऊ / नोटिस बोर्ड पर चरपा हेतु।
- प्रभारी, कम्प्यूटर सेल, कार्यालय प्रमुख सचिव, चिकित्सा स्वास्थ्य एवं परिवार कल्याण, उ०प्र० शासन।

अपर निदेशक (भण्डार)